


ANALYZED BY:

Anresco Laboratories
1375 Van Dyke Avenue,
San Francisco, CA 94124
DEA# PA0202945

CUSTOMER:

Laurelcrest Labs
1270 NE Alpha Drive
McMinnville, OR 97128



SAMPLE INFORMATION

Sample No.: 1412497
Product Name: CBG ISOLATE
Matrix: Concentrate (Isolate)
Lot #: BK-26-134

Date Collected: 05/15/2026
Date Received: 05/15/2026
Date Reported: 05/21/2026

TEST SUMMARY

Cannabinoid Profile: ✔ Tested

Residual Solvent Screen: ✔ Pass

Cannabinoid Profile

05/21/2026

Method: MF-CHEM-15
Instrument: Liquid Chromatography Diode Array Detector (LC-DAD)
Limit of Detection 0.3333 mg/g
Limit of Quantitation 1.0000 mg/g

Cannabinoid	mg/g	%
Δ8-THC	ND	ND
Δ9-THC	<LOQ	<LOQ
Δ9-THCA	ND	ND
THCV	ND	ND
THCVA	ND	ND
CBD	ND	ND
CBDA	ND	ND
CBC	1.4	0.14
CBCA	ND	ND
CBDV	ND	ND
CBG	995.0	99.50
CBGA	ND	ND
CBN	ND	ND
Total THC	<LOQ	<LOQ
Total CBD	ND	ND
Total Cannabinoids	996.3	99.63
Sum of Cannabinoids	996.3	99.63

Total THC = Δ8-THC + Δ9-THC + (0.877 * THCA)
Total CBD = CBD + (0.877 * CBDA)
Total Cannabinoids = Σ (neutral cannabinoids) + [0.877 * Σ (acidic cannabinoids)]

The result of this sample is confirmed with a retest.

Comment(s):

The reported result for delta-9 THC represents the quantitation of a peak that elutes in the retention time window of the delta-9 THC certified reference standard. However, it is inconclusive due to the interference of an unknown compound. LC-MS/MS analysis is recommended for a more conclusive identification and quantitation.

Residual Solvent Screen ✔ Pass

05/21/2026

Method: MF-CHEM-32

Regulatory Standard* DCC

Instrument: Gas Chromatography Mass Spectrometry (GC/MS)

Analyte	LOD/LOQ (µg/g)	Findings (µg/g)	Limit (µg/g)*	Status
1,2-Dichloroethane	0.5/0.5	ND	1	Pass
Acetone	57/200	ND	5000	Pass
Acetonitrile	56/200	ND	410	Pass
Benzene	0.5/0.5	ND	1	Pass
n-Butane	45/200	ND	5000	Pass
Chloroform	0.5/0.5	ND	1	Pass
Ethanol	37/200	ND	5000	Pass
Ethyl acetate	38/200	ND	5000	Pass
Ethyl ether	37/200	ND	5000	Pass
Ethylene oxide	0.1/0.5	ND	1	Pass
n-Heptane	135/200	560.00	5000	Pass
n-Hexane	49/200	ND	290	Pass
Isopropyl alcohol	57/200	ND	5000	Pass
Methanol	37/200	ND	3000	Pass
Methylene chloride	0.1/0.5	ND	1	Pass
n-Pentane	37/200	ND	5000	Pass
Propane	72/200	ND	5000	Pass
Toluene	49/200	ND	890	Pass
Total xylenes (ortho-, meta-, para-)	58/200	ND	2170	Pass
Trichloroethylene	0.5/0.5	ND	1	Pass

ND = None Detected
LOD = Limit of Detection
LOQ = Limit of Quantitation

Reported by




Vu Lam
Lab Co Director



Scan to verify